

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0511M

System: Skyrad PIRD1

Site: Manus

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
PIR	31301F3	Skyrad	11/4/2004	-0.75 K0 .31270 K1 0.999 K2 (-3.52) K3	na	11/15/2005	03:00	04:15

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 ⁻⁶ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
PIR	31312F3	Skyrad	11/15/2005	-4.06 K0 .29104 K1 1.005 K2 (-2.572) K3	na	X

Replacement completed by: Troy Culkan

NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other comments: